	Application No.	Applicant(s)	
Notice of Allowability	10/645,810	CIARROCCA, SCOTT ANDREW	
	Examiner	Art Unit	
	Peter J. Vrettakos	3739	
The MAILING DATE of this communication appears on the cover sheet with the correspondence address All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS. This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.			
1. 🔀 This communication is responsive to <u>Amendment 10-31-06</u> .			
2. X The allowed claim(s) is/are <u>1,3-11,13,15-18,20 and 21</u> .			
3.			
Attachment(s) 1. Notice of References Cited (PTO-892) 2. Notice of Draftperson's Patent Drawing Review (PTO-948) 3. Information Disclosure Statements (PTO/SB/08), Paper No./Mail Date 4. Examiner's Comment Regarding Requirement for Deposit of Biological Material	5. ☐ Notice of Informal P 6. ☐ Interview Summary Paper No./Mail Dat 7. ☐ Examiner's Amendr 8. ☒ Examiner's Stateme 9. ☐ Other	(PTO-413), te ment/Comment	owance

REASONS FOR ALLOWANCE

The following is an examiner's statement of reasons for allowance: Independent claims 1, 8, 13, 15 and 21 disclose a bipolar electrosurgical instrument with a converting element with two positions one in contact with an active electrode and a second position in contact with a return electrode all with both of the positions affecting the effective surface area of the active electrode relative to the return electrode. This articulation is depicted in figures 4a and 4b. The converting element is element 42, the active electrode is element 40 and the return electrode is 44. The prior art neglects to disclose this dynamic configuration described in each of independent claims 1, 8, 13, 15 and 21.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Peter J. Vrettakos whose telephone number is 571-272-4775. The examiner can normally be reached on M-F 9-6.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda C. Dvorak can be reached on 571-272-4764. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Pete Vrettakos December 27, 2006

> ROY D. GIBSON PRIMARY EXAMINER